



TEST CERTIFICATE
ISSUED BY SIRA TEST & CERTIFICATION LIMITED

**TEST FOR THE INGRESS PROTECTION OF
AN nVISION HAND HELD CALIBRATOR**

Supplier: Crystal Engineering Corporation
708 Fiero Lane
Suite 9
San Luis Obispo
California 93401
USA

Model or Type Identification: nVision hand held calibrator

Standard: BS EN 60529:1992 Incorporating Amendments Nos 1 and 2

Deviations from Standard: None

ST&C Test Procedure: LOP 220

ST&C Test Reports: 09/0179 and N52A19587A

Samples Delivery Date: 9 March 2009

Tests Conducted Between: 23 to 26 March 2009

This certificate refers to the performance of the test samples when tested against the agreed programme. It does not imply that any other samples or products necessarily comply with the requirements of the test programme.

Sira Test & Certification Limited being a UKAS accredited Test House in accordance with ISO/IEC 17025 has tested the above nVision hand held calibrator, and has found it to comply with the requirements of the Ingress Protection Code: IP 67.

S P Cork
Laboratory Manager

Dated 14 April 2009

This certificate may only be reproduced in its entirety and without change

Page 1 of 1

Certificate No: Sira 52A19587ALab
Form 6240, Issue 8

a Volvere plc group company